



21ST 
WORLD
STERILIZATION
CONGRESS

17 / 20 NOVEMBER
2021

**CICG, GENEVA,
SWITZERLAND**





Schweizerische Gesellschaft für Sterilgutversorgung
Société Suisse de Stérilisation Hospitalière
Società Svizzera di Sterilizzazione Ospedaliera

Increasing patient safety by validating a reproducible method for reprocessing ultrasound probes

Part 2

Jalel Ben Mesmia
CSSD Strategic Manager
University Hospital Schleswig-Holstein
Germany

<p>European Union level</p> 	<p>European Medical Devices Regulation</p> <ul style="list-style-type: none"> • MDR (Medical Device Regulation) <ul style="list-style-type: none"> • (General provisions)
<p>National level</p> 	<ul style="list-style-type: none"> • MPDG (Medical Devices Implementation Act) <ul style="list-style-type: none"> • (Penal Code) • MPBetreibV (Medical Devices Operator Ordinance) <ul style="list-style-type: none"> • (Detailed requirements) • (IfSG) Infection Protection Act <ul style="list-style-type: none"> • (Prevention and control of infectious diseases)
<p>Sub-legislative level</p>	<ul style="list-style-type: none"> • Standards (DIN-EN-ISO) • Guidelines: <ul style="list-style-type: none"> • Guideline of the Commission for Hospital Hygiene and Infection Prevention (KRINKO) at the Robert Koch Institute (RKI) and the Federal Institute for Drugs and Medical Devices (BfArM) "DGKH/RKI/BfArM". • Scientific societies (DGSV/SGSV/ÖGSV) <ul style="list-style-type: none"> • Detailed instructions for practical use • Regulations/Rules <ul style="list-style-type: none"> • UVV (Accident Prevention Regulations) • BGR (Employer's Liability Insurance Association Rules) • BGV (Employer's Liability Insurance Association Regulations)

- **MDR**

- **11. infection and microbial contamination**
 - 11.1 Devices and their manufacturing processes shall be designed in such a way as to eliminate or to reduce as far as possible the risk of infection to patients, users and, where applicable, other persons.

 - 11.5 Devices labelled as 'sterile' shall be processed, manufactured, packaged and sterilised by means of appropriate, validated methods.

Prevention measures!

- Protection against Infection Act (IfSG) § 23 Para. 3

Requirements	■ (3) The heads of the following facilities shall ensure that the <u>necessary measures are taken</u> in accordance with the state of the art in medical science,
Goal	to <u>prevent healthcare-associated infections</u> and avoid the <u>spread of pathogens, especially those with resistance:</u> <ul style="list-style-type: none">■ Hospitals,■ Outpatient surgery facilities,■ Day Hospitals,■ Maternity facilities,■ Medical practices, dental practices,■ ...

- Medical Devices Operator Ordinance §8.1
 - (1) The reprocessing of medical devices intended to be used disinfected or sterile shall be carried out, **taking into account the manufacturer's instructions**, using **suitable validated procedures** in such a way that the success of these procedures is **ensured and traceable**
 - Requirements**
 - and the **safety and health of** patients, users or third parties is **not endangered**.
 - Goal**

- Risk assessment and classification
 - Diagnostics
 - Semi-Critical
 - Intraoperative
 - Critical
- Grouping
 - Group A
- Characteristics
 - Sensitive to liquids and heat!



- Wipe disinfection procedure with:
 - Disposable cleaning wipes + Disposable disinfecting wipes
- Semi-automatic cleaning and disinfection process with „validatable high-level disinfection process“

❖ 20.11.2020 Publication of a technical explanation by the Robert Koch Institute with the following key statements

- *"...a validability of the final disinfection of semi-critical medical devices with wipes is currently ... not given"*
 - **Manual wipe disinfection cannot be validated**
- *"The RKI is not aware of any guideline or standard that could serve as an adequate basis for ensuring this requirement."*
 - **Lack of validation guidelines**
- *"German law requires on-site validation as part of the mandatory validation requirements for reprocessing semi-critical medical devices."*
 - **The Robert Koch Institute points out that a validated process must be generated by the operator for both Automated and manual reprocessing of medical devices.**

- Wipe disinfection procedure with:
 - Disposable cleaning wipes + Disposable disinfecting wipes


Wipe disinfection as a procedure?

- Semi-automatic cleaning and disinfection processes with “validatable high-level disinfection process”

Currently the only validatable method or alternative to wipe disinfection.

Semi-automatic process

Sequence of a semi-automatic cleaning and disinfection process




Validatable sub-process? Validatable sub-process?

Is the validation of the automatic disinfection device or the "disinfection process" sufficient to consider the entire process as a validated process?

No!

■ Cleaning and disinfection procedures

- Purely automated process: automated cleaning and automated disinfection process
- Purely manual process: manual cleaning and manual disinfection process
- Semi-automatic process: manual cleaning process and automated disinfection process



Validation Guidelines

Automation:


➤ "Guideline of DGKH, DGSV and AKI for the validation and routine monitoring of automated cleaning and thermal disinfection processes for medical devices".

- DIN EN ISO 14971, DIN EN 15883 and DIN EN ISO 17664

Manual:

➤ "Guideline for validation of manual cleaning and manual chemical disinfection of medical devices".

- DIN EN ISO 14971, DIN EN 14885 and DIN EN ISO 17664




Validation principles of a semi-automatic cleaning and disinfection process

Automated:

- "Guideline of DGKH, DGSV and AKI for the validation and routine monitoring of automated cleaning and thermal disinfection processes for medical devices".
- DIN EN ISO 14971, DIN EN 15883 and DIN EN ISO 17664

Manual:

- "Guideline for validation of manual cleaning and manual chemical disinfection of medical devices".
- DIN EN ISO 14971, DIN EN 14885 and DIN EN ISO 17664



HYBETA GmbH • Neuvogt 20 • 43147 Münster
 Universitätsklinikum Schleswig-Holstein ADR
 - Campus Kiel -
 Schwanenweg 20
 24105 Kiel



HYBETA GmbH
 Labor Münster
 Neuvogt 20
 43147 Münster
 Fax: +49 (0)201 2851-118
 Fax: +49 (0)201 2851-129
 labor-m@hybeta.com

Prüfbericht zur Auftragsnummer A2021016392 vom 27.09.2021

Auftrag

Auftragsnummer	A2021016392	Kundennummer	11857
Prüfung	Validierung manuelle Aufbereitung - Desinfektionsprüfung		
Probenstelle	Endoskopie		
Probennehmer	HYBETA GmbH	Probenahmedatum	21.09.2021
Hersteller/Gerätetyp		Seriennummer	
Art der Aufbereitung	maschinel, manuell	Desinfektionsmittel	
Referenznummer			

Prüfung

Eingang	23.09.2021	Prüfbeginn	23.09.2021	Prüfende	27.09.2021	Freigabe	27.09.2021
---------	------------	------------	------------	----------	------------	----------	------------

Prüfergebnisse

Probennummer	Material	Testler
A2021016392-001		
Lokalisation		
Parameter	Wert	Einheit
Keime	kein Keimwachstum	

Probennummer	Material	Testler
A2021016392-002		
Lokalisation		
Parameter	Wert	Einheit
Keime	kein Keimwachstum	

Probennummer	Material	Testler
A2021016392-003		
Lokalisation		
Parameter	Wert	Einheit
Keime	kein Keimwachstum	

HYBETA GmbH
 Neuvogt 20
 43147 Münster

Geschäftsführer
 Dr. Frank Wille
 Dr. Cornelia Liber

StZ der Gesellschaft
 Antigenhof Münster IHD 8398
 USt-IDN: DE239419234

Seite 1 von 2

HYBETA GmbH • Neuvogt 20 • 43147 Münster
 Universitätsklinikum Schleswig-Holstein ADR
 - Campus Kiel -
 Schwanenweg 20
 24105 Kiel



HYBETA GmbH
 Labor Münster
 Neuvogt 20
 43147 Münster
 Fax: +49 (0)201 2851-118
 Fax: +49 (0)201 2851-129
 labor-m@hybeta.com

Prüfbericht zur Auftragsnummer A2021016394 vom 23.09.2021

Auftrag

Auftragsnummer	A2021016394	Kundennummer	11857
Prüfung	Validierung manuelle Aufbereitung		
Probennehmer	HYBETA GmbH	Probenahmedatum	21.09.2021
Hersteller/Gerätetyp		Seriennummer	
Probennummer			
Probenstelle	Endoskopie		

Prüfung

Eingang	23.09.2021	Prüfbeginn	23.09.2021	Prüfende	23.09.2021	Freigabe	23.09.2021
---------	------------	------------	------------	----------	------------	----------	------------

Prüfergebnisse

Probennummer	Material	Eluat
A2021016394-001		
Beschreibung		
R1		
Parameter	Wert	Einheit
Restprotein	< 5	µg

Probennummer	Material	Eluat
A2021016394-002		
Beschreibung		
R2		
Parameter	Wert	Einheit
Restprotein	< 5	µg

Probennummer	Material	Eluat
A2021016394-003		
Beschreibung		
R3		
Parameter	Wert	Einheit
Restprotein	< 5	µg

HYBETA GmbH
 Neuvogt 20
 43147 Münster

Geschäftsführer
 Dr. Frank Wille
 Dr. Cornelia Liber

StZ der Gesellschaft
 Antigenhof Münster IHD 8398
 USt-IDN: DE239419234

Seite 1 von 2

Pr fbericht zur Auftragsnummer A2021016394 vom 23.09.2021

Auftrag

Auftragsnummer	A2021016394	Kundennummer	11857
Pr�fung	Validierung manuelle Aufbereitung		
Probenehmer	- HYBETA GmbH	Probenahmedatum	21.09.2021
Hersteller/Ger�tetyt		Seriennummer	
Referenznummer			
Probenstelle	Endoskopie		

Pr fung

Eingang	23.09.2021	Pr�fbeginn	23.09.2021	Pr�fende	23.09.2021	Freigabe	23.09.2021
---------	------------	------------	------------	----------	------------	----------	------------

Pr fergebnisse

Probenummer	A2021016394-001	Material	Eluat	Wert	Einheit
Beschreibung	R1				
Parameter					
Restprotein				< 5	�g

Probenummer	A2021016394-002	Material	Eluat	Wert	Einheit
Beschreibung	R2				
Parameter					
Restprotein				< 5	�g

Probenummer	A2021016394-003	Material	Eluat	Wert	Einheit
Beschreibung	R3				
Parameter					
Restprotein				< 5	�g

Pr fbericht zur Auftragsnummer A2021016392 vom 27.09.2021

Auftrag

Auftragsnummer	A2021016392	Kundennummer	11857
Pr�fung	Validierung manuelle Aufbereitung - Desinfektionspr�fung		
Probenstelle	Endoskopie		
Probenehmer	HYBETA GmbH	Probenahmedatum	21.09.2021
Hersteller/Ger�tetyt		Seriennummer	
Art der Aufbereitung	maschinell, manuell	Desinfektionsmittel	
Referenznummer			

Pr fung


Eingang	23.09.2021	Pr�fbeginn	23.09.2021	Pr�fende	27.09.2021	Freigabe	27.09.2021
---------	------------	------------	------------	----------	------------	----------	------------

Pr fergebnisse

Probenummer	A2021016392-001	Material	Tupfer	Wert	Einheit
Lokalisation					
Parameter					
Keime				kein Keimwachstum	

Probenummer	A2021016392-002	Material	Tupfer	Wert	Einheit
Lokalisation					
Parameter					
Keime				kein Keimwachstum	

Probenummer	A2021016392-003	Material	Tupfer	Wert	Einheit
Lokalisation					
Parameter					
Keime				kein Keimwachstum	



Thank you for your attention!

Speakers:

Dr. Maja DECIUS

Mr Jalel BEN MESMIA

University Hospital Schleswig-Holstein

Emails:

Maja.Decius@uksh.de

Jalel.BenMesmia@uksh.de

